

SPONSOR 12993



Triangle Fluid Controls
399 College Street E.
Belleville, ON K8N 5S7
Canada

REPORT DATE

September 26, 2016

CERTIFICATE
CLASS VI - 121°C PLASTIC TESTS
(USP 39 <88>, REV. 5/2016)

Name: Durlon 9000N
Physical Description: Solid
Lot Number: 5755
Sample Code: Not provided by Sponsor
Part Number: TN15-060-060
Expiration Date: Not provided by Sponsor
Sterilization Data: Non-sterile; Autoclaved at Pacific BioLabs at 121°C for approximately 30 minutes
Storage Conditions: Room Temperature
Date Received: July 22, 2016

Systemic Injection Test in CD-1 Mice
Pacific BioLabs Report No.: 16G0212R-X02

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits
Pacific BioLabs Report No.: 16G0212R-X03

Implantation Test in New Zealand White Rabbits
Pacific BioLabs Report No.: 16G0212R-X04

The test article met the requirements for Plastic Class VI - 121°C testing as per USP <88> guidelines.

Multiple Study Project Coordinator: Zuzana Karjala, Ph.D., RLATg, Senior Scientist

9/26/16

X

I approve the content of this document.
Signed by: Zuzana Karjala

:ahc